

**Konformitätserklärung
Declaration of Conformity**

Wir

We

**B. Braun Melsungen AG
Carl-Braun-Str. 1
34212 Melsungen
Deutschland/Germany
SRN DE-MF-000000201**erklären in eigener Verantwortung,
dass das Produkt**Ecoflac® Mix**

Transfer-Kappe für sterile Flüssigkeiten

Basis UDI-DI: 403923900000027222
(Artikelnummern siehe Anlage I)mit den Anforderungen der Medizinprodukte
Verordnung (EU) 2017/745 übereinstimmt**Konformitätsbewertungsverfahren**
nach Anhang IX
der oben genannten Verordnung**Klassifizierung**
gemäß Anhang VIII der oben genannten
Verordnung
Klasse I steril**Benannte Stelle**
TÜV SÜD Product Service GmbH
Kennnummer 0123**Gültig bis**
gemäß gültigem EU Zertifikat
(G11 012974 0626)hereby declare in our sole responsibility
that the product**Ecoflac® Mix**

Transfer cap for sterile fluids

Basis UDI-DI: 403923900000027222
(article numbers see attachment I)is in conformity with the requirements of the
Medical Device Regulation (EU) 2017/745**Conformity Assessment Procedure**
according to annex IX
of the Regulation named above**Classification**
according to annex VIII of the Regulation named
above
Class I sterile**Notified Body**
TÜV SÜD Product Service GmbH
Identification number 0123**Valid until**
according to our valid EU Certificate
(G11 012974 0626)

Anlage I / Attachment I

Basic UDI-DI 403923900000027222

Art.-Nr. / Art. No. Produktname / Product name

16401N Ecoflac® Mix

Klasse / Class

I steril / I sterile

Document amendment information

Version	Description of the changes
1.0	Initial Version under 2017/745 MDR

Title: Declaration of Conformity - 214-005-MDR - Ecoflac Mix Initiator: Susanne ? Olbricht

This document is signed electronically in compliance with the B. Braun electronic signature policies and procedures by following persons:

UserName: Olbricht, Susanne (olbrsude)
Title: HC-RA-DE08E Regulatory Affairs Manager IV-Systems
Date: Friday, 24 May 2024, 13:57 W. Europe Daylight Time
Meaning: Document signed as Author
=====

UserName: Brand, Thomas (brantode)
Title: HC-QM-DE08 Vice President QM for non-active Medical Devices
Date: Friday, 24 May 2024, 16:29 W. Europe Daylight Time
Meaning: Approve Document
=====

UserName: Seidel, Stefan (seidstde)
Title: Head of Regulatory Affairs CoE Infusion & Pain Therapy
Date: Friday, 24 May 2024, 17:34 W. Europe Daylight Time
Meaning: Approve Document
=====